



NOCHM-4

Multi-Port Measurement Chamber

INSTRUCTION MANUAL

Serial No. _____

012414

www.wpiinc.com

World Precision Instruments

CONTENTS

ABOUT THIS MANUAL 1

INTRODUCTION 1

 Features of the Multi-Port Measurement Chamber..... 1

 Parts List 2

 Unpacking 2

INSTRUMENT DESCRIPTION 2

 Set-Up..... 3

OPERATING INSTRUCTIONS..... 3

 Installing Side Fitting Plug in the Side Fitting Cap..... 3

 Room Temperature Use 3

 Mounting a Probe..... 5

 Using with Circulating Water Bath 6

INSTRUMENT MAINTENANCE..... 6

 Cleaning 6

 Sterilization..... 6

ACCESSORIES..... 7

SPECIFICATIONS..... 8

WARRANTY 9

 Claims and Returns 9

 Repairs 9

ABOUT THIS MANUAL

The following symbols are used in this guide:



This symbol indicates a **CAUTION**. Cautions warn against actions that can cause damage to equipment. Please read these carefully.



This symbol indicates a **WARNING**. Warnings alert you to actions that can cause personal injury or pose a physical threat. Please read these carefully.

NOTES and TIPS contain helpful information.

INTRODUCTION

The Multi-Port Measurement Chamber is a temperature stabilized closed chamber for measurements of nitric oxide, oxygen, hydrogen peroxide, hydrogen sulfide, ions, and other species in cell culture. It is also ideal for the calibration of NO electrodes at controlled temperature. The measurement of NO and other reactive gases will be underestimated in stirred solutions if the solution is allowed to equilibrate with air. In the case of NO, accelerated decomposition occurs as the result of diffusion of NO from the solution into the gas phase and the reaction of NO with oxygen. This reaction with oxygen makes a significant and variable contribution to NO decomposition, and hence accuracy of measurement, at concentrations of NO between 0.1-5 μM . These problems can now be eliminated with the use of the Multi-Port Measurement Chamber, which minimizes the headspace of the experimental solution used. In addition to NO a number of other analytes can be measured with WPI's line of 2-mm sensors.

Features of the Multi-Port Measurement Chamber

- Closed chamber design greatly reduces the surface area of the experimental solution exposed to air.
- The individual port openings are matched to WPI's 2 mm tip diameter **ISO-NOP**, **ISO-H2S**, **ISO-HPO**, **ISO-OXY-2** and **OXELP** electrodes, **Kwik-Tip™** ion-selective electrodes and **DRIREF-2** and **SDR2** reference electrodes for the measurement of calcium, potassium, hydrogen and TPP.



CAUTION: Not recommended for low (<1% concentration) measurements.

- Three optional side ports can be used for additional electrode insertion, permitting simultaneous measurements of up to three analytes.
- Two openings in the top fitting cap for injection of reagents into the chamber.
- Operating temperature range of 4-40°C controlled via an external circulating bath.
- Ideal for NO electrode calibration across a range of temperatures.

Parts List

- (1) **NOCHM-4** NO Chamber
- (1) Top Fitting Cap
- (3) Side Fitting Caps
- (1) Package of 10 Gaskets (WPI #G04)
- (3) Side Fitting Cap Plugs

Unpacking

Upon receipt of this product, make a thorough inspection of the contents and check for possible damage. Missing cartons or obvious damage to cartons should be noted on the delivery receipt before signing. Concealed loss or damage should be reported at once to the carrier and an inspection requested. Please see "Claims and Returns" on page 9.


Returns: Do not return any goods to WPI without obtaining prior approval and instructions from our Service Department. Goods returned (unauthorized) by collect freight may be refused. If a return shipment is necessary, use the original container. If the original container is not available, use a suitable substitute that is rigid and of adequate size. Wrap the instrument in paper or plastic surrounded with at least 100 mm (four inches) of shock absorbing material. Please see "Claims and Returns" on page 9.

INSTRUMENT DESCRIPTION

The description below refers to Figure 1.

The NOCHM consists of a Chamber (**A**) and a Top Fitting Cap (**B**), through which a 2-mm electrode can be inserted (see Fig. 1). Three optional use Side Ports (**C**) are also provided through which three additional electrodes can be inserted for the simultaneous measurement of up to three analytes. The inner volume of the chamber (and hence sample volume) can be continuously adjusted in volume from 1.0mL to 3.0mL and is suitable for most cell suspension experiments and electrode calibrations.

The NO Chamber can be temperature-controlled by circulating water through the outer jacket of the chamber using an appropriate heating/cooling circulator. The circulator bath is connected to the NO Chamber through Inlet (**D**) and Outlet (**E**) Connectors.



The Top Fitting Cap (**B**) has 2 openings (**F**) for reagent addition to the chamber. WPI's MicroFil™ syringe needles are suitable for this purpose.

Set-Up

No assembly is required before use.

OPERATING INSTRUCTIONS

! **CAUTION:** During all manipulations involving the electrodes, avoid contact with the electrode tips. The membrane covering the tip is fragile and damage could occur upon contact with materials such as the rubber gasket or the plastic of the chamber.

See Instrument Maintenance for cleaning procedures. It is recommended that, at a minimum, the unit be flushed well with deionized water before use.

Installing Side Fitting Plug in the Side Fitting Cap

The Side Fitting Plug (WPI #30018) effectively prevents leakage and closes a Side Fitting Cap (WPI #30014) that is not needed for experimental measurements. Simply insert the plug into the Side Fitting Cap as shown in Figure 2. The plug can easily be pulled out when an electrode needs to be inserted into the cap.



Fig. 2—Installing the Side Fitting Plug.

Room Temperature Use




Fig. 3—Diagram of Side Fitting Electrode Cap with electrode and gasket in place.

Step 1: Side Port

! **CAUTION:** The Side Fitting Caps (WPI # 33013) has a gasket at its inner tip, even when using the Side Fitting Plug (WPI # 33018). The gasket is required to effectively seal off the port. Use the **G04** gasket when replacing it.

1. If the Side Port is not to be used, ensure that the Side Fitting Cap (WPI # 33013) (with properly installed Side Fitting Plug (WPI # 33018) and gasket) is in place.
2. To place an electrode in the Side Port, remove the Side Fitting Cap from the side port. If the Side Fitting Cap has a Side Fitting Plug, pull out the plug.
3. Insert the electrode through the opening in the Side Fitting Electrode Cap. Once inserted, install a gasket on the electrode sleeve, about 5mm from the end (Fig. 3).

TIP: wet the gasket first.

4. Carefully place the Side Fitting Cap (WPI # **33013**), with the electrode, into the Side Port. Adjust the position of the electrode so that it extends into the chamber solution and tighten the Side Fitting Cap.

CAUTION: Moving the electrode in or out after tightening the Cap, although possible, may cause damage to the membrane at the electrode tip. Electrode tips are fragile. To avoid damaging the membrane tip, when disassembling the Side Fitting Cap and electrode from the chamber, first loosen and then remove the Cap with the electrode still inserted. Carefully remove the gasket from the electrode and then remove the electrode from the Cap.

Step 2: Top Port:

1. Slide the Top Fitting Cap (B) out of the Chamber (A).
 2. Fill the chamber with the experimental solution (1-3mL); add a stirrer bar if applicable.
 3. Taking care to minimize contact with the electrode tip, fit the electrode through one of the gaskets provided.
- TIP:** wet the gasket first.

CAUTION: To decrease the possibility of damaging the electrode membrane while reinstalling the gasket, it is recommended that the gasket be left on the top electrode between uses.

4. Slide the electrode into the opening on the outer side of the Top Fitting Cap (WPI # **30012**) just until it protrudes through the cap to the inner chamber side.
5. Replace the Fitting Cap into the top port.
6. Position the bottom of the Top Fitting Cap at the surface of the experimental solution. Ensure that at least one millimeter of the electrode is in the solution. Push the electrode further down into the Cap as necessary. The gasket holds the electrode in place and seals the chamber. (Fig. 4)
7. The two additional openings on the Top Fitting Cap are provided for injection of reagents into the chamber.

NOTE: Although both the top and side ports can be used for either NO or oxygen measurement, the top port cannot be plugged and therefore must be used when only one electrode is employed.




Fig. 4—Top Fitting Cap with electrode in place. Base of cap (see arrow) should touch the surface of the solution in the chamber.

Mounting a Probe

In Fig. 5 (below) the probe parts are laid out in order of assembly.




Fig. 5–Disassembled probe

1. Slide the gasket on the mounting tube as shown in Fig. 6.
2. Insert the probe tip carefully through the acrylic knob and then gently slide the mounting tube with the gasket past the membrane. The tube does not have to go all the way onto the membrane sleeve just far enough to clear the upper edges of the membrane.



Fig. 6–The mounting tube with the gasket slides over the membrane sleeve.

3. Slide the gasket off the mounting tube onto the membrane sleeve (Fig. 7).



Fig. 7–Slide the gasket onto the membrane sleeve.

4. Slide off the mounting tube (Fig. 8). Position the gasket near the end of the membrane and insert it into the chamber until the gasket seats. Check the position of the probe depth into the chamber and then tighten the acrylic screw until the gasket compresses and seals.



Fig. 8–Remove the mounting tube.

Using with Circulating Water Bath



CAUTION: Exercise caution when connecting a heating/cooling circulator bath to the Multi-port Measurement Chamber. Failure to properly secure the tubing to the inlet/outlet connectors of the Multi-port Measurement Chamber may result in flooding.

1. To use the Multi-port Measurement Chamber at temperatures above or below room temperature but within the range of 4-40° C, connect an appropriate heating/cooling circulator bath to the inlet/outlet connectors (E and D in Fig. 1) using 1/4-in. tubing.
2. Refer to the manual of your circulator bath on how to adjust the temperature.
3. When starting the circulator for the first time turn the chamber upside down to eliminate any trapped bubbles in the internal paths.
4. Once the circulating solution has reached the desired temperature, proceed with the experiment as described in the room temperature section above.

NOTE: Once solution has been added to the chamber, additional temperature equilibration of the solution may be required, depending on its temperature at the time it was added.

NOTE: You will have to determine the correct temperature setting for the circulating bath in order to achieve the desired temperature of the sample in the Multi-port Measurement Chamber. They may not be the same.

INSTRUMENT MAINTENANCE

Cleaning

After use, rinse the chamber compartment and the Fitting Caps with water. To clean the compartment of protein residue use an enzymatic detergent like ENZOL (WPI #ENZOL).

Sterilization



CAUTION: DO NOT AUTOCLAVE the Multi-Port Measurement Chamber.

Ethylene oxide can be used to sterilize as well as Cidex Plus (WPI #7364). Although alcohol is acceptable, over time it may cause damage to the acrylamide unit.

ACCESSORIES

WPI Part #	Description
33012	Top Fitting Cap
33013	Side Fitting Cap
33018	Side Fitting Plug
500787	Haake DC10-P5/U Circulating Water Bath (115v)
500788	Haake DC10-P5/U Circulating Water Bath (230v)
500789	Haake DC10-P5/U Circulating Water Bath (100v)
7364	Cidex Plus
7357	Nitrite Standard Solution
DRIREF-2	Dri-Ref, 2 mm diameter electrode
Enzol	Enzymatic Detergent
G04-100	Gasket, 2 mm ID, package of 100
GSNO-50	GSNO, 50 mg
GSNO-100	GSNO, 100 mg
GSNO-500	GSNO, 500 mg
ISO-HPO-2	Hydrogen Peroxide electrode
ISO-H2S	Hydrogen Sulfide electrode
ISO-NOP	ISO-NOP electrode (2 mm tip dia)
ISO-OXY-2	Oxygen electrode (2 mm tip dia)
KWIKCAL-2	Holder and 3 Calcium electrodes
KWIKH-2	Holder and 3 Hydrogen electrodes
KWIKPOT-2	Holder and 3 Potassium electrodes
KWIKTPP-2	Holder and 3 TPP (tetraphenylphosphonium) electrodes
MF34G-5	MicroFil™, 34 ga., 67 mm Long (pkg. of 5)
MF28G-5	MicroFil™, 28 ga., 97 mm Long (pkg. of 5)
MF28G67-5	MicroFil™, 28 ga., 67 mm Long (pkg. of 5)
OXELP	Oxygen electrode (2 mm tip dia)
SDR2	SUPER-Dri-Ref, 2 mm diameter electrode
SNAP50	SNAP, 50 mg
SNAP100	SNAP, 100 mg
SNAP500	SNAP, 500 mg

SPECIFICATIONS

Volume of Sample	1-3mL
Sample Injection Ports	2 (top)
Electrode Ports	4
Electrode Compatibility	Nitric Oxide Electrode ISO-NOP
Oxygen Electrode	OXELP (Not recommended for low (<1%) concentration measurements)
	ISO-OXY-2 (Not recommended for low (<1%) concentration measurements)
Oxygen Electrode	
Hydrogen Peroxide Electrode	ISO-HPO-2
Hydrogen Sulfide Electrode	ISO-H2S
Calcium Electrode	KWIKCAL-2
Hydrogen Electrode	KWIKH-2
Potassium Electrode	KWIKPOT -2
TPP (tetraphenylphosphonium) Electrode	KWIKTPP-2
Dri-Ref Electrode	DRIREF-2
SUPER-Dri-Ref Electrode	SDR2
Temperature Range of Circulating Water	4-40 °C
Water inlet and outlet require 1/4-in. ID tubing	

WARRANTY

WPI (World Precision Instruments, Inc.) warrants to the original purchaser that this equipment, including its components and parts, shall be free from defects in material and workmanship for a period of 30 days* from the date of receipt. WPI's obligation under this warranty shall be limited to repair or replacement, at WPI's option, of the equipment or defective components or parts upon receipt thereof f.o.b. WPI, Sarasota, Florida U.S.A. Return of a repaired instrument shall be f.o.b. Sarasota.

The above warranty is contingent upon normal usage and does not cover products which have been modified without WPI's approval or which have been subjected to unusual physical or electrical stress or on which the original identification marks have been removed or altered. The above warranty will not apply if adjustment, repair or parts replacement is required because of accident, neglect, misuse, failure of electric power, air conditioning, humidity control, or causes other than normal and ordinary usage.

To the extent that any of its equipment is furnished by a manufacturer other than WPI, the foregoing warranty shall be applicable only to the extent of the warranty furnished by such other manufacturer. This warranty will not apply to appearance terms, such as knobs, handles, dials or the like.

WPI makes no warranty of any kind, express or implied or statutory, including without limitation any warranties of merchantability and/or fitness for a particular purpose. WPI shall not be liable for any damages, whether direct, indirect, special or consequential arising from a failure of this product to operate in the manner desired by the user. WPI shall not be liable for any damage to data or property that may be caused directly or indirectly by use of this product.

Claims and Returns

Inspect all shipments upon receipt. Missing cartons or obvious damage to cartons should be noted on the delivery receipt before signing. Concealed loss or damage should be reported at once to the carrier and an inspection requested. All claims for shortage or damage must be made within ten (10) days after receipt of shipment. Claims for lost shipments must be made within thirty (30) days of receipt of invoice or other notification of shipment. Please save damaged or pilfered cartons until claim is settled. In some instances, photographic documentation may be required. Some items are time-sensitive; WPI assumes no extended warranty or any liability for use beyond the date specified on the container

Do not return any goods to us without obtaining prior approval and instructions from our Returns Department. Goods returned (unauthorized) by collect freight may be refused. Goods accepted for restocking will be exchanged or credited to your WPI account. Goods returned which were ordered by customers in error are subject to a 25% restocking charge. Equipment which was built as a special order cannot be returned.

Repairs

Contact our Customer Service Department for assistance in the repair of apparatus. Do not return goods until instructions have been received. Returned items must be securely packed to prevent further damage in transit. The Customer is responsible for paying shipping expenses, including adequate insurance on all items returned for repairs. Identification of the item(s) by model number, name, as well as complete description of the difficulties experienced should be written on the repair purchase order and on a tag attached to the item.

** Electrodes, batteries and other consumable parts are warranted for 30 days only from the date on which the customer receives these items.*



World Precision Instruments, Inc.

USA

International Trade Center, 175 Sarasota Center Blvd., Sarasota FL 34240-9258

Tel: 941-371-1003 • Fax: 941-377-5428 • E-mail: sales@wpiinc.com

UK

1 Hunting Gate, Hitchin, Hertfordshire SG4 0TJ

Tel: 44 (0)1462 424700 • Fax: 44 (0)1462 424701 • E-mail: wpiuk@wpi-europe.com

Germany

Zossener Str. 55, 10961 Berlin

Tel: 030-6188845 • Fax: 030-6188670 • E-mail: wpide@wpi-europe.com

China & Hong Kong

WPI Shanghai Trading Co., Ltd.

Rm 29a, No8 Dongfang Rd., Pudong District, Shanghai, 200120 PR China

Tel: +86 21 6888 5517 • E-mail: chinasales@china.wpiinc.com

Internet

www.wpiinc.com • www.wpi-europe.com • www.wpiinc.cn